



Pulse oximeter measurement error of oxygen saturation in patients with SARS-CoV-2 infection stratified by smoking status

Colin J. Crooks $0^{1,2,3}$, Joe West $0^{2,3,4,5}$, Joanne R. Morling $0^{2,3,4}$, Mark Simmonds, Irene Juurlink 0^{3} , Steve Briggs 0^{3} , Simon Cruickshank 0^{3} , Susan Hammond-Pears $0^{3,5}$, Dominick Shaw $0^{3,6}$, Timothy R. Card $0^{2,3,4}$ and Andrew W. Fogarty $0^{2,3,4}$

¹Nottingham Digestive Diseases Centre, School of Medicine, University of Nottingham, Nottingham, UK. ²NIHR Nottingham Biomedical Research Centre (BRC), Nottingham University Hospitals NHS Trust and the University of Nottingham, Nottingham, UK. ³Nottingham University Hospitals NHS Trust, Nottingham, UK. ⁴Lifespan and Population Health, School of Medicine, University of Nottingham, Nottingham, UK. ⁵East Midlands Academic Health Science Network, University of Nottingham, Nottingham, UK. ⁶NIHR Biomedical Respiratory Research Centre University of Nottingham, Nottingham, UK.

Corresponding author: Andrew W. Fogarty (andrew.fogarty@nottingham.ac.uk)



Shareable abstract (@ERSpublications)

There is substantial measurement error in pulse oximetry readings of oxygen saturation below 90% which is not due to smoking status https://bit.ly/3RunKtL

Cite this article as: Crooks CJ, West J, Morling JR, et al. Pulse oximeter measurement error of oxygen saturation in patients with SARS-CoV-2 infection stratified by smoking status. Eur Respir J 2022; 60: 2201190 [DOI: 10.1183/13993003.01190-2022].

This single-page version can be shared freely online.

Copyright $\mbox{\em @The authors 2022}.$

This version is distributed under the terms of the Creative Commons Attribution Licence 4.0.

Received: 26 April 2022 Accepted: 26 Sept 2022 To the Editor:

Pulse oximeters provide a non-invasive measurement of oxygen saturation and are now routinely used to assess patients, inform medical decision-making and monitor subsequent clinical status.



